North American Diagnostics
NADIAGNOSTICS.COM
+1 201-357-0230
807 Brazos St, Austin, Texas 78701
HPLC Purity | Mass Content | Mass Identification

Endotoxin Test Report

Test ID	NAD-0037990
Client Name	ProSupply Peptides | pspeptides.com
Date Ordered	October 28th, 2026

Product	Acetic Acid	**Appearance**	Clear Colorless Liquid
Mass Reported	0.6%; 99.4%	**Vial Count**	1x | 3mL Vial
Notes	1x 3mL vial (0.6% acetic acid + distilled water)	**Batch ID**	PSP-10172026AA3

Endotoxin Testing Performed By LAL Assay Using Kinetic Chromogenic Method

Test Method	Endotoxin Level
LAL Kinetic Chromogenic	**< 0.5 EU/mL**

Overall Status
PASSED — Below 0.5 EU/mL Limit

Mark Brownstein, Analysis Administration Lead